

# CHARGER



Platform Scale

**R260**

User's Manual



---

Please read this manual carefully before operating your device and retain it for future reference.

IN-00261 [REV001] 07/19

# CONTENTS

**INTRODUCTION ..... 2**  
    Safety Precautions ..... 2

**INSTALLATION ..... 3**  
    Package Contents..... 3  
    Parts Description..... 4  
    Selecting the Location ..... 4  
    Power Supply ..... 4  
    Battery Power ..... 5

**OPERATION ..... 5**  
    Keypad and Display Description ..... 5  
    Status Indicators ..... 6  
    Keypad Description ..... 6  
    Turning Scale On/Off..... 6  
    Zero Operation..... 6  
    Changing Units of Measure..... 7

**MAINTENANCE ..... 7**  
    Cleaning ..... 7  
    Troubleshooting..... 8  
    Service Information ..... 8

**TECHNICAL DATA ..... 9**  
    Specifications..... 9  
    Disposal..... 10

**ERROR AND STATUS MESSAGE ..... 10**  
    Error Message..... 10

**WARRANTY ..... 11**

## **INTRODUCTION**

This manual provides detailed instructions for the installation work, operation and maintenance instructions of the CHARDER R260 Scale, please read this manual thoroughly before installing and operating the scale.

### Safety Precautions

Please follow these safety precautions:

- Verify AC Adapter input voltage matches the local AC power supply.
- Do not insert liquids or foreign object.
- Do not operate the scale in hostile environments.
- Do not drop loads or leave permanent load on the platform. This may damage the measuring system.
- Do not place the scale upside down on the pan.
- Do not open the scale. Never attempt to make repairs yourself because this could be dangerous. The warranty is void if this stipulation is ignored. The scale may only be opened by authorized personnel.
- Always disconnect the device from the power supply before installing, servicing, cleaning or performing maintenance.
- Check the power cable regularly for damage. If it is damaged, immediately disconnect the device from the power supply.

## **INSTALLATION**

### Package Contents

- R260 Scale
- AC Power Adaptor
- User's Manual




————— R260 Scale



————— AC Power Adaptor




————— User's Manual

Parts DescriptionSelecting the Location

Select a firm, vibration-free and horizontal location for the scale. Avoid locations with direct sunlight, excessive air current, heat sources, or rapid temperature fluctuations.

Power Supply

Connect the AC Adaptor to the AC mains supply. Connect the plug to the AC jack on the bottom of the scale as shown in the illustration below. The power outlet should be close to the scale and easily accessible.



## Battery Power

The scale will automatically switch to battery operation if there is a power failure or the power cord is removed to provide uninterrupted scale operation.

**Note:** When low battery  is displayed, the scale will shut off after 30 seconds.



### **CAUTION:**

Risk of explosion can occur if replaced with the wrong type or connected improperly.

Do not use the battery for any purpose other than those described in this manual.

Do not disassemble, repair or otherwise manipulate the battery.  
Obey the local laws and regulations on disposing the battery.

## **OPERATION**

### Keypad and Display Description



Status Indicators

ICON	Description
	Stability display
	Zero setting activated
	Negative sign
NET	Display net weight
Hold	Display hold function
kg / lb	Currently chosen weight unit

Keypad Description

KEY	DESCRIPTION	FUNCTION
	POWER ON / OFF	- Turn on and off display
	UNIT	- Change the weighing units to alternate units of measurement
	TARE	- Call up TARE function
	HOLD	- Call up HOLD function

Turning Scale On/Off

To turn the scale on, press the **[ON / OFF]** button scale will display **88888** and then enters the active weighing mode.

To turn the scale off, press the **[ON / OFF]** button to turn it off.

Zero Operation

When power on the scale and if the weight on the scale is within the initial zero tolerance, the scale will display zero automatically.

### Changing Units of Measure

In weighing mode, use **[UNIT]** to display the next available measuring unit. Repeat process to scroll through all units.

### Performing TARE function

Place an empty container on the scale and press [TARE]. The display will show a net value of "0".

To clear the Tare value, press [TARE] with the pan empty.

### Performing HOLD feature

Press [HOLD] and a triangle will appear above NET triangle. And load object on the platform. Once the weight is stable, the scale locks the weight automatically. Even if you remove the object from the platform. Press [HOLD] again to release HOLD feature.

## **MAINTENANCE**

### Cleaning

Before cleaning, disconnect the power plug from power outlet and remove batteries from the appliance to power off the unit.

- Clean with a damp cloth and a mild soap or similar agents if necessary, and buffed dry with.
- Do not use any acids, alkalis or strong active solvents to clean the housing or control panels.

Troubleshooting

<b>SYMPTOM</b>	<b>POSSIBLE CAUSE</b>	<b>REMEDY</b>
Scale does not turn on	Battery discharged	Replace the battery or use adapter
	Power cord not plugged in properly	Check power cord connection and plugged in properly into the power outlet.
	Other failure	Have the scale serviced by Charder authorized service center.
Poor accuracy	Unstable environment	Move the scale to suitable location, refer to the manual <i>Selecting the Location</i> .
	Improper calibration	Perform calibration.

**TECHNICAL DATA**Specifications

<b>MODEL</b>	<b>R260</b>				
Capacity	6 kg / 12 lb	15 kg / 30 lb	30 kg / 60 lb	60 kg / 120 lb	150 kg / 300 lb
Graduation	2 g / 0.005 lb	5 g / 0.01 lb	5 g < 15 kg > 10 g 0.01 lb < 30 lb > 0.02 lb	10 g / 0.02 lb	20 g / 0.05 lb
Power Requirements	100 V - 240 V DC 15V AC adapter 4-AA batteries				
Weighing Units	kg / lb				
Display	1" in LCD with backlight				
Display Type	LCD with backlit				
Key Function	4 Keys ON,OFF / TARE / UNIT / HOLD				
Operating Temperature and Humidity	5°C to 35° C / 41°F to 95°F 15% - 85% RH				
Application Mode	Weighing , Hold , Tare				
Construction	SUS 304 Stainless Steel platform				
Platform Size	315 x 305 x 55 mm				
Weight	4 kg				

Disposal

In conformance with the European Directive 2002/96 EC on Waste Electrical and Electronic Equipment (WEEE), this device may not be disposed of with domestic waste. This also applies to countries outside the EU, according to their specific requirements.

Please dispose of this product in accordance with local regulations at the collecting point specified for electrical and electronic equipment. For more detailed information about recycling this product, please contact the local distributor from which you purchased this device.

**ERROR AND STATUS MESSAGE**Error Message

ERROR MESSAGE	DESCRIPTION	REMEDY
	Low battery indication	Connect the power supply or change new batteries.
	Zero value is too high or too low when power on	Empty load plate before power up and check if scale is placed correctly. Recalibrate the scale.
	Zero value is too high when power on	Empty load plate before power up. Recalibrate the scale.
	AD board is malfunctioned	Check possible disconnect wire or change AD board.
	AD board is malfunctioned when entering calibration mode	Change AD board.

## **WARRANTY**

Charder scales are warranted against defects in material and workmanship for a period of one(1) year from the date of purchase. Please retain your receipt as proof of purchase. Warranty expiration on products returned without proof of purchase will be determined from the manufacturing date code or the date of shipment to the authorized dealer. Coverage may be invalidated as this date is previous to purchase date. No other express or implied warranty is given by Charder.

Freight damage on returns is not covered under warranty.

This warranty does not apply for following reasons, the product has been damaged by accident or misuse, parts and products of a consumable nature, exposed to radioactive or corrosive materials, has foreign material penetrating to the inside of the product, or as a result of service or modification by other than Charder.